

**दर अनुबंध बिड दस्तावेज़ / Rate Contract Bid Document**

[This Rate Contract Bid is governed by General Terms and Conditions of Rate Contract.](#)

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	18-03-2026 14:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	18-03-2026 14:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Labour And Employment
विभाग का नाम/Department Name	Na
संगठन का नाम/Organisation Name	Employees State Insurance Corporation (esic)
कार्यालय का नाम/Office Name	Headquarters
वस्तु श्रेणी /Item Category	Neonatal Ventilators (Q2)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	1944 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Additional Doc 4 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)

बिड विवरण/Bid Details	
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / <b>Minimum number of bids required to disable automatic bid extension</b>	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / <b>Number of days for which Bid would be auto-extended</b>	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / <b>Number of Auto Extension count</b>	1
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
दर अनुबंध और शीर्षक / <b>Rate Contract &amp; Title</b>	Yes ; Rate Contract for Neonatal Ventilator
दर अनुबंध की वैधता अवधि / <b>Validity Period Of Rate Contract</b>	12 Months
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय / <b>Time allowed for Technical Clarifications during technical evaluation</b>	5 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/ <b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
अनुमानित बिड मूल्य / <b>Estimated Bid Value</b>	48600000
मूल्यांकन पद्धति/ <b>Evaluation Method</b>	Item wise consignee wise evaluation

#### ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
-----------------------------	---------------------

#### Discount Slab Range

#### Neonatal Ventilators

Slab 1	1 - and above Unit
--------	--------------------

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD should be in favour of Beneficiary, wherever it is applicable.

**लाभार्थी /Beneficiary :**

Director General, ESIC

Room No. 321, III Floor, ESIC Hqrs. Office, Panchdeep Bhawan, C.I.G. Road, New Delhi - 110002

(Esic Fund Account No. 1)

बोली विभाजन लागू नहीं किया गया/Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

**Pre Bid Detail(s)**

मूल्य भिन्नता खंड दस्तावेज/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
02-03-2026 11:00:00	Main Conference Hall, 5th Floor, ESIC Hqrs. Office, Panchdeep Bhawan, C.I.G. Road, New Delhi - 110002

मूल्यांकन विधि(मदवार मूल्यांकन विधि) / **Evaluation Method & Consignee Location** ( Item Wise Consignee Wise Evaluation Method )

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The details of item-consignee combination covered under each schedule are as under:

मूल्यांकन अनुसूचियां / <b>Evaluation Schedules</b>	वस्तु/श्रेणी / <b>Item/Category</b>	<b>Location of Consignee</b>
Schedule 1 ( CENTRAL DELHI )	Neonatal Ventilators	CENTRAL DELHI

## Neonatal Ventilators

### तकनीकी विशिष्टियाँ / **Technical Specifications**

\* जेम केटेगरी विशिष्टि के अनुसार / [As per GeM Category Specification](#)

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
GENERAL	Product Description	Neonatal Ventilator
PERFORMANCE PARAMETERS	<b>Patient Type</b>	Neonatal, Premature
	<b>Tidal Volume Range In ml</b>	2 to 300
	<b>Trigger Sensitivity Mechanism</b>	Pressure, Flow, Pressure and Flow
	<b>Flow Pattern/Waveform Adjustment</b>	Yes
	<b>Sigh Breath Function</b>	Yes, No
	<b>Leak Compensation Available</b>	Yes
	<b>Auto 100% increase O2 Button</b>	Yes
	<b>Inspiratory Pause Available</b>	Yes
	<b>Expiratory Hold Available</b>	Yes, No
	<b>Suction Control Option Available</b>	Yes, No
	<b>Nasal CPAP Option Available</b>	Yes
	<b>Auto PEEP Manoeuvre</b>	Yes, No
	<b>Control Panel Lock</b>	Yes, No
	<b>Facility for double lung ventilation</b>	Yes, No
	<b>Integrated Air Compressor</b>	Yes
VENTILATION MODES	<b>PSV(Pressure supported ventilation)</b>	Yes
	<b>P-CMV(Inspiration pressure control)</b>	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>PC-AC With Expiratory Minimal volume</b>	Yes, No
	<b>SIMV(Synchronized intermittent Mandatory Ventilation) with PCMV</b>	Yes
	<b>ITT - Inspiratory Time Termination</b>	Yes, No
	<b>CPAP:Continuos Positive Airway pressure Ventilation CPAP With backup Ventilation</b>	Yes
	<b>High Frequency Oscillation Ventilator(HFOV)</b>	Yes, No
PATIENT ASSESSMENT TOOLS	<b>Capnography/Co2 Monitoring</b>	Yes, No
	<b>Volume Guarantee</b>	Yes, No
	<b>APRV</b>	Yes, No
	<b>Oxygen Therapy</b>	Yes, No
ADDITIONAL CAPABILITIES	<b>ETCO2 Inbuilt available</b>	Yes, No
HUMIDIFIER	<b>Disposable Neonatal breathing circuit (dual heated wire)</b>	20, 50, 80, 100, NA
	<b>Auto Feed Humidification chamber Available</b>	Yes
	<b>Disposable Neonatal breathing circuit with single heated wire provided for inspiratory Limb with Water trap in expiratory limb available</b>	20, 50, 80, 100, NA
ADDITIONAL PARAMETERS	<b>Output Ports Type</b>	HDMI, Ethernet, RJ 45, USB
	<b>Remote Alarm Display Port</b>	Yes
	<b>Reporting (Vent alarms and Patient Status)</b>	Yes
	<b>View Reports on Display</b>	Yes
	<b>Save Data to USB</b>	Yes
	<b>Send Data Via Network</b>	Yes, No
	<b>Display Type</b>	LCD/LED/TFT
	<b>Display Size in Inches</b>	? 10, ? 12, ? 15 Or higher
	<b>Can be used For Patient Transport</b>	Yes, No
ACCESSORIES	<b>Nasal CPAP Kit</b>	10, 15, 20

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	<b>Number of Disposable flow sensor</b>	20, 50, 100, NA if Reusable
	<b>Number of Reusable flow sensor</b>	3, 6, NA if Disposable
	<b>Trolley with ventilator with supporting arm from same manufacturer</b>	Yes
	<b>Autoclavable exhalation valve/expiratory cassette</b>	2
	<b>Test Lung Available</b>	Yes
POWER REQUIREMENTS	<b>Operating time of battery in min</b>	30, 60, 120
WARRANTY	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	<b>5.0 (year)</b>

### Special terms and conditions-Version:1 effective from 15-04-2024 for category Neonatal Ventilators

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling

- price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
  6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
  8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
  10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service

centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

## क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

### 1. **Rate Contract**

The Purchase orders placed through the rate of this bid shall be placed by . The order shall be placed by buyer separately based on the requirement.

### 2. **Rate Contract**

Tax and Duties Toll Tax/ Entry Tax/ Octroi Duty & Local Taxes. No separate payment would be made by the Buyer for Toll Taxes, Entry Taxes, Octroi duty and local Taxes, if any. The Seller should cater for these Taxes/ duties as part of Basic Rate quoted in the Bid.

### 3. **Rate Contract**

**Demurrage charges** In case the rejected items are not lifted by the firm within 48 hrs, the demurrage charges at the rate of 0.5% of total contract value will be charged per day. In case the items are not lifted within a month, the same will be destroyed by the station board of officers and no claim will be admitted. Demurrage charges. In case the rejected items are not lifted by the firm within 48 hrs, the demurrage charges at the rate of 0.5% of total contract value will be charged per day. In case the items are not lifted within a month, the same will be destroyed by the station board of officers and no claim will be admitted.

### 4. **Rate Contract**

**Packing Material** The items will be supplied by the successful bidder in its original packing material and the packing material will not be returned. Weight of packing material will not be included in quantity supplied and only net weight of the items will be counted. The packing should be of standardized weights of

appropriate size. Item will not be accepted in non-standardized weights. Packing Material. The items will be supplied by the successful bidder in its original packing material and the packing material will not be returned. Weight of packing material will not be included in quantity supplied and only net weight of the items will be counted. The packing should be of standardized weights of appropriate size. Item will not be accepted in non-standardized weights.

## 5. Rate Contract

**Non return of Hard Disk:** As per Buyer organization's Security Policy, Faulty Hard Disk of Servers/Desktop Computers/ Laptops etc. will not be returned back to the OEM/supplier against warranty replacement.

## 6. Rate Contract

**End User Certificate:** Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.

## 7. Rate Contract

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

## 8. Rate Contract

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

## 9. Rate Contract

The Bidder / OEM {themselves or through reseller(s)}, should have executed project for supply and installation / commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:

- (i) Single order of at least 35% of estimated bid value; or
- (ii) Two orders of at least 20% each of estimated bid value; or
- (iii) Three orders of at least 15% each of estimated bid value.

Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion

## 10. Rate Contract

Malicious Code Certificate:

The seller should upload following certificate in the bid:-

(a) This is to certify that the Hardware and the Software being offered, as part of the contract, does not contain Embedded Malicious code that would activate procedures to :-

- (i) Inhibit the desired and designed function of the equipment.
- (ii) Cause physical damage to the user or equipment during the exploitation.
- (iii) Tap information resident or transient in the equipment/network.

(b) The firm will be considered to be in breach of the procurement contract, in case physical damage, loss of information or infringements related to copyright and Intellectual Property Right (IPRs) are caused due to activation of any such malicious code in embedded software.

## 11. Rate Contract

**OEM Turn Over Criteria:** The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution

shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

## 12. **Rate Contract**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

## 13. **Rate Contract**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

## 14. **Rate Contract**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

## 15. **Rate Contract**

AVAILABILITY OF OFFICE OF SERVICE PROVIDER: An office of the Service Provider must be located in the state of Consignee. DOCUMENTARY EVIDENCE TO BE SUBMITTED.

## 16. **Rate Contract**

ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.

## 17. **Rate Contract**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

## 18. **Rate Contract**

**Non- Disclosure:** The OEM/bidder has to sign a Non-Disclosure agreement with the buyer for Software license products like Anti-Virus solutions etc.

## 19. **Rate Contract**

**NET WORTH:** Net Worth of the OEM should be positive as per the last audited financial statement.

## 20. **Rate Contract**

IT equipment shall be IPv6 ready from day one.

## 21. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## अस्वीकरण/**Disclaimer**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses

on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

**All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into force by the Government of India.**

**For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.**

**The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:**

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

**All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.**

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

**---धन्यवाद/Thank You---**